


Layout VT 4002



- Entry port NW 25 mm installed in basic equipment
L1
- Entry port
L2 further installation position l.h. side (option)
R1 further installation position r.h. side (option)
- 1 Main switch
- 2 Minicontrol
- 3 Independent adjustable temperature limiter t_{\min}/t_{\max}
- 4 not assigned
- 5 Power supply, length of cable approx. 5 m
- 6 Door with window
- 7 Connector panel

+ For narrow entry openings components (e.g. hinges) may be screwed off to minimize width.